



## Haloethers EPA Method 611/8110

Column: **MEGA-5 - 0.53mm, 1.50 $\mu$ m, 15m**

Catalog Code: S-5-053-150-15

### Conditions:

Injection: Split 250°C, 0.2 $\mu$ L, 1:50 Split Ratio.

Sample: 2000 $\mu$ g/mL each in Acetone.

Detector: FID 300°C.

Oven Temp.: 100°C, 15°C/min, 300°C.

Carrier Gas: Helium, 20 kPa.

### Peaks:

1. Bis (2-chloroethyl) ether
2. Bis (2-chloroisopropyl) ether
3. Bis (2-chloroethoxy) methane
4. 4-chlorophenyl phenyl ether
5. 4-bromophenyl ether

